




GHRP-6 HPLC Purity Test Report

Product Name :GHRP-6
 Batch No. :P170606G6
 Column :Xbridge BEH130 C18, 4.6*250mm, 5um
 Solvent A :0.1%Trifluoroacetic in 100% Acetonitrile
 Solvent B :0.1%Trifluoroacetic in 100% Water
 Gradient : A B
 0.0min 26% 74%
 25.0min 51% 49%
 25.1min 100% 0%
 30.0min Stop
 Flow rate :1.0ml/min
 Wavelength :220nm



Rank	Time	Quantity	Area	Height
1	12.004	0.5482	101161	16545
2	12.490	99.3231	18329944	1483179
3	12.958	0.1287	23760	5712
Total		100	18454865	1505436